

## Deutsche Akkreditierungsstelle GmbH

**Entrusted according to Section 8 subsection 1 AkkStelleG in connection with Section 1 subsection 1 AkkStelleGBV**

Signatory to the Multilateral Agreements of EA, ILAC and IAF for Mutual Recognition

# Accreditation



The Deutsche Akkreditierungsstelle GmbH attests that the calibration laboratory

**VACUUBRAND GmbH & CO KG**  
**Alfred-Zippe-Straße 4, 97877 Wertheim**

is competent under the terms of DIN EN ISO/IEC 17025:2018 to carry out calibrations in the following fields:

### **Mechanical quantities**

- **Vacuum**

The accreditation certificate shall only apply in connection with the notice of accreditation of 10.12.2020 with the accreditation number D-K-15154-01. It comprises the cover sheet, the reverse side of the cover sheet and the following annex with a total of 2 pages.

Registration number of the certificate: **D-K-15154-01-00**

Berlin,  
10.12.2020

Dr Heike Manke  
Head of Division

Translation issued:  
14.06.2021

Head of Division



*The certificate together with the annex reflects the status as indicated by the date of issue.  
The current status of any given scope of accreditation may be found respectively in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH <https://www.dakks.de/en/content/accredited-bodies-dakks>.*

This document is a translation. The definitive version is the original German accreditation certificate.

See notes overleaf.

# Deutsche Akkreditierungsstelle GmbH

Office Berlin  
Spittelmarkt 10  
10117 Berlin

Office Frankfurt am Main  
Europa-Allee 52  
60327 Frankfurt am Main

Office Braunschweig  
Bundesallee 100  
38116 Braunschweig

The publication of extracts of the accreditation certificate is subject to the prior written approval by Deutsche Akkreditierungsstelle GmbH (DAkKS). Exempted is the unchanged form of separate disseminations of the cover sheet by the conformity assessment body mentioned overleaf.

No impression shall be made that the accreditation also extends to fields beyond the scope of accreditation attested by DAkKS.

The accreditation was granted pursuant to the Act on the Accreditation Body (AkkStelleG) of 31 July 2009 (Federal Law Gazette I p. 2625) and the Regulation (EC) No 765/2008 of the European Parliament and of the Council of 9 July 2008 setting out the requirements for accreditation and market surveillance relating to the marketing of products (Official Journal of the European Union L 218 of 9 July 2008, p. 30). DAkKS is a signatory to the Multilateral Agreements for Mutual Recognition of the European co-operation for Accreditation (EA), International Accreditation Forum (IAF) and International Laboratory Accreditation Cooperation (ILAC). The signatories to these agreements recognise each other's accreditations.

The up-to-date state of membership can be retrieved from the following websites:

EA: [www.european-accreditation.org](http://www.european-accreditation.org)

ILAC: [www.ilac.org](http://www.ilac.org)

IAF: [www.iaf.nu](http://www.iaf.nu)

## Deutsche Akkreditierungsstelle GmbH

### Annex to the Accreditation Certificate D-K-15154-01-00 according to DIN EN ISO/IEC 17025:2018

Valid from: 10.12.2020

Date of issue 10.12.2020

Holder of certificate:

**VACUUBRAND GmbH & CO KG**  
**Alfred-Zippe-Straße 4, 97877 Wertheim**

Calibration in the fields:

**Mechanical quantities**  
- Vacuum

Within the measurands / calibration items marked with \*) , the calibration laboratory is permitted, without being required to inform and obtain prior approval from DAkkS, to use calibration standards or equivalent calibration procedures listed here with different issue dates. The calibration laboratory maintains a current list of all calibration standards / equivalent calibration procedures within the flexible scope of accreditation. In house calibration procedures are excluded from the flexible scope.

*The management system requirements of DIN EN ISO/IEC 17025 are written in the language relevant to the operations of calibration laboratories. Laboratories that conform to the requirements of this standard, operate generally in accordance with the principles of DIN EN ISO 9001.*

*The certificate together with the annex reflects the status as indicated by the date of issue.  
The current status of any given scope of accreditation may be found respectively in the database of accredited bodies of Deutsche Akkreditierungsstelle GmbH <https://www.dakks.de/en/content/accredited-bodies-dakks>.*

Abbreviations used: see last page

**Page 1 of 2**

**This document is a translation. The definitive version is the original German annex to the accreditation certificate.**

**Annex to the accreditation certificate D-K-15154-01-00**

**Permanent Laboratory**

**Calibration and Measurement Capabilities (CMC)**

Measurement quantity / Calibration item	Range	Measurement conditions / procedure	Expanded uncertainty of measurement <sup>1)</sup>	Remarks
<b>Vacuum*</b> Absolute pressure p / vacuum gauges	1,0·10 <sup>-3</sup> mbar to < 2,0·10 <sup>-3</sup> mbar	DIN ISO 3567:2015 ISO 27893:2011 ISO 19685:2017 ISO 20146:2019	1,00 %	Pressure medium: Nitrogen (N <sub>2</sub> )
	2,0·10 <sup>-3</sup> mbar to < 1,0·10 <sup>-2</sup> mbar		0,80 %	
	1,0·10 <sup>-2</sup> mbar to < 2,0 mbar		0,75 %	
	2,0 mbar to < 3,0 mbar		0,50 %	
	3,0 mbar to < 6,0 mbar		0,40 %	
	6,0 mbar to < 20,0 mbar		0,35 %	
	20,0 mbar to < 1100 mbar		0,25 %	
	1100 mbar to ≤ 1300 mbar	VACUU CAL 1300 Vers. A:2020	0,25 %	

**Abbreviations used:**

CMC	Calibration and measurement capabilities
DIN	Deutsches Institut für Normung e.V.
ISO	International Organization for Standardization
VACUU CAL	In house calibration procedure of Vacuubrand

<sup>1)</sup> The expanded uncertainties according to EA-4/02 M:2013 are part of CMC and are the best measurement uncertainties within accreditation. They have a coverage probability of approximately 95 % and have a coverage factor of  $k = 2$  unless stated otherwise. Uncertainties without unit are relative uncertainties referring to the measurement value unless stated otherwise.